

fluticasone propionate inhalation powder

Overview of how to use Aermony RespiClick®

INSTRUCTIONS FOR PATIENTS

For more information and an instructional video, visit

www.respiclick.ca

Scan this QR code or type the URL into your browser.



Log in with the DIN on the outside of your Aermony RespiClick® package or with this number: 02467895.

To use: Open. Inhale. Close.



OPEN the cap all the way until you feel and hear a "click"—once the cap clicks, one dose is ready to be inhaled.

Make sure you are holding the inhaler upright. Before going to step 2, hold the inhaler away from your mouth and breathe out through your mouth as much air as you can and as is comfortable.

- Never breathe out into the inhaler mouthpiece.
- · Do not open the green cap unless you are taking a dose.



Inhale.

INHALE quickly and deeply through the mouth, until lungs feel completely full of air.

Your lips should form a good seal around the mouthpiece, but the vent above the mouthpiece should not be blocked by the lips or fingers. Remove the inhaler from your mouth and hold your breath for about 10 seconds or for as long as you comfortably can.

 Do not take an extra dose from the inhaler even if you do not taste or feel the medicine.



CLOSE the cap after inhaling so that the inhaler will be ready for the next dose.

Rinse your mouth with water after taking your dose. Spit out the water. Do not swallow it.

Throw away your inhaler 30 days after you have opened the foil pouch or when the counter reads "0," whichever comes first.

Please read the medication leaflet that came with your inhaler for more information about how to take Aermony RespiClick®.







Aermony RespiClick® is a registered trademark of Ivax International B.V., a wholly-owned subsidiary of Teva Pharmaceutical Industries Ltd., and is used under licence.

TEVA and the design version thereof are registered trademarks

of Teva Pharmaceutical Industries Ltd. and are used under licence.

© 2021 Teva Canada Innovation G.P. – S.E.N.C., Montreal, Quebec H2Z 1S8